

Anti-Ustekinumab Neutralizing Antibody ELISA kit

Summary

Catalog No.	KAD84051
Alternative Names	CNTO 1275, TT-20, CAS: 815610-63-0
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
Detection method	Colorimetric
Sample type	Plasma, Serum
Assay type	Quantitative
Range	15.625-1,000 ng/ml
Recovery	80-120%
Shipping	2-8 °C
Note	For Research Use Only.

Precision

CV < 20%

Data Image
