

# SARS-CoV-2 Surrogate Virus Neutralization Test Kit

## Summary

---

<b>Catalog No.</b>	KAV99908
<b>Alternative Names</b>	Spike glycoprotein, S glycoprotein, Spike protein S1, S, Receptor-Binding Domain, RBD, Neutralization, SARS-CoV-2
<b>Stability and Storage</b>	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
<b>Detection method</b>	Colorimetric
<b>Sample type</b>	Plasma, Serum
<b>Assay type</b>	Quantitative
<b>Sensitivity</b>	78.1 ng/mL
<b>Range</b>	0.156 - 10 µg/mL
<b>Recovery</b>	80-120%
<b>Shipping</b>	2-8 °C
<b>Note</b>	For Research Use Only.

## Precision

---

CV<15%

## Data Image

---