

SARS-CoV-2 Surrogate Virus Neutralization Test Kit (B.1617.2/L452R, T478K)

Summary

Catalog No.	KAV99902
Alternative Names	Spike glycoprotein, S glycoprotein, Spike protein S1, S, Receptor-Binding Domain, RBD, India-202010, Delta, B.1617.2, G/452R.V3, 21A/S:478K, Neutralization, SARS-CoV-2
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
Detection method	Colorimetric
Sample type	Plasma, Serum
Assay type	Quantitative
Sensitivity	78.1 ng/mL
Range	0.156 - 10 µg/mL
Recovery	80-120%
Shipping	2-8 °C
Note	For Research Use Only.

Precision

CV<15%

Data Image
