

# Treponema pallidum TPP17 ELISA Kit

## Summary

---

<b>Catalog No.</b>	KXX06701
<b>Specificity</b>	Treponema pallidum TPP17
<b>Stability and Storage</b>	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition.
<b>Detection method</b>	Colorimetric
<b>Sample type</b>	Plasma, Serum
<b>Assay type</b>	Quantitative
<b>Sensitivity</b>	0.65 ng/mL
<b>Range</b>	1.56 - 100 ng/mL
<b>Recovery</b>	80-120%
<b>Shipping</b>	2-8 °C
<b>Note</b>	For research use only.

## Precision

---

CV<15%

## Data Image

---