

11α -hydroxyprogesterone/11- α -OH-P ELISA Kit

Summary

Catalog No. KSK61602

> HYDROXYPROGESTERONE,11A-; DELTA4-PREGNEN-11ALPHA-OL-3,20-DIONE;(8S,9S,10R,11R,13S,14S,17S)-17-ACETYL-11-HYDROXY-10,13-DIMETHYL-1,2,6,7,8,9,10,11,12,13,14,15,16,17-TETRADECAHYDRO-

Alternative Names CYCLOPENTA[A]PHENANTHREN-3-ONE;11A-HYDROXY-4-PREGNENE-

> 3,20-DIONE;11A-HYDROXYPROGESTERONE;11ALPHA-HYDROXY-4-PREGNENE-3,20-DIONE;11ALPHA-HYDROXYPREGN-4-ENE-3,20-

DIONE;11ALPHA-HYDROXYPROGESTERONE

The stability of ELISA kit is determined by the loss rate of activity. The loss

Stability and Storage rate of this kit is less than 10% prior to the expiration date under

appropriate storage condition.

Detection method Colorimetric

Sample type Plasma, Serum, Urine

Assay type Quantitative (Competitive)

Sensitivity 3 ng/mL

Range 4.7 ng/mL - 300 ng/mL

Recovery 80-120%

Shipping 2-8 °C

Note For research use only.

Precision





CV

Data Image