

Suptavumab ELISA Kit

Summary

Catalog No. KDV02803

H1H3592P3, REGN 2222, SAR-438584, SAR438584, REGN2222, CAS: **Alternative Names**

1629615-23-1

The stability of ELISA kit is determined by the loss rate of activity. The loss

Stability and Storage rate of this kit is less than 10% prior to the expiration date under

appropriate storage condition.

Detection method Colorimetric

Sample type Plasma, Serum

Assay type Quantitative

Sensitivity 0.156 µg/ml

Range $0.31-5 \mu g/mL$

Recovery 80-120%

Shipping 2-8 ℃

Note For Research Use Only.

Precision

CV<20%

Data Image